

Medispo®

Premium

HBM Protections

WE MAKE THE BEST, WE CARE THE MOST

HBM Protections



Guilin HBM Health Protections, Inc. stands as a distinguished leader in the global production of specialist surgical gloves. With a legacy spanning over two decades, we have consistently maintained our position among the top three producers worldwide. Our unwavering commitment to excellence has earned us accreditation to ISO EN13485 quality management systems and MDR status, accompanied by coveted CE Mark and 510K approval for both the European Union and the United States of America.


ENVIRONMENTAL COMMITMENT:

- Clean Energy: We harness clean energy sources, utilizing high-pressure steam and electricity at the point of use.
- Chlorine Elimination: Our production process is entirely chlorine-free, aligning with our commitment to environmental sustainability.
- Renewable Energy: We collaborate closely with local government authorities to source energy from renewable methods, reducing our carbon footprint.
- Radioactive Waste Mitigation: Our in-house E-Beam sterilization technique eliminates the concerns associated with radioactive waste generated by gamma ray sterilization.

ENVIRONMENTAL STEWARDSHIP: Our corporate headquarters are nestled in a pristine natural environment, and we uphold the highest standards of pollution control. We continuously innovate to minimize our ecological impact, with a steadfast commitment to achieving zero carbon emissions at the point of manufacture. Our energy supply, driven by high-pressure, high-temperature steam and electricity, underscores our commitment to clean and sustainable practices. Our aspirational target is to attain carbon neutrality by the year 2035.

CORPORATE SOCIAL RESPONSIBILITY (CSR): At Guilin HBM, we take pride in our role as responsible corporate citizens. Every member of our workforce resides within the local community, and we do not employ immigrant labour. We are dedicated to providing a hygienic and secure work environment, offering competitive compensation to attract and retain talent while fostering a positive relationship with our community.

OPERATIONAL EXCELLENCE: Efficiency and innovation are at the core of our operations. Our cutting-edge production platform incorporates computerized control and employs Artificial Intelligence for predictive quality analytics, enabling proactive issue prevention. This technological prowess positions our state-of-the-art facility among the most efficient in the global landscape.



WORLD LEADING QUALITY ASSURANCE SYSTEMS, LABORATORY TESTING AND CORPORATE SOCIAL RESPONSIBILITY CREDENTIALS


- **Quality Management Systems**
  - ISO 9001 / ISO EN 13485 / MDSAP Certified
- **MDR EU Certificate MDR 747912**
  - Notified Body BSI Group 2797
- **PPE Certificate**
  - Notified Body SATRA 2777
- **CE Mark & FDA 510K** Clearance for EU & USA markets
- **ISO 14001**, Environmental Management
- **ISO 45001**, Occupational Health & Safety
- Test Laboratories and notified bodies
  - BSI, SATRA, TUV, SGS
  - ARDL – Chemotherapy drug permeation
- Corporate Social Responsibility (CSR)
  - V-TRUST Certification

Leading the Way in Sustainable Surgical Gloves Technology

HBM Protections

WORLD LEADING QUALITY ASSURANCE SYSTEMS

- FDA 510K
- MDR EU Quality Management System Certificate MDR 747912, Notified Body BSI Group The Netherlands B.V. 2797
- ISO 9001 / ISO 13485 / MDSAP Certified
- ISO 14001, Environmental Management
- ISO 45001, Occupational Health & Safety
- PPE Certificate, Notified Body BSI Group The Netherlands B.V. 2797





















HBM Protections

WORLD CLASS LABORATORY TESTING

- **EN455-1/2/3/4 - tested by PSB / Singapore**
- **Chemo Drug Permeation Test ASTM D6978 - by ARDL / USA**
- **Physical property comparison testing by internationally regarded laboratory demonstrates industry leading product performance**
- **PPE - tested by SATRA**
- **AQL-Exceeds all international standards**



Test Report No. 719125302S-EEC21-WBH
dated 08 Feb 2021

SUBJECT:
Testing of Surgical Gloves submitted by Guilin HBM Health Protections, Inc. on 29 Jan 2021.

TESTED FOR:
Guilin HBM Health Protections, Inc.
No.1-2, Shujing East Road,
Economic and Technological Development Area,
Guilin, China

TEST DATE:
02 Feb 2021

DESCRIPTION OF SAMPLES:


S/N	Product Description	Brand/ Model	Size	Model No.	Lot No.	Expiry Date	Sample Received (pairs)	Manufacturer
1	Surgical Gloves, Sterile, Latex, Powder-Free	Medline	5.5	MSG7255	GUIL20W47	2022-10	11	Guilin HBM Health Protections, Inc.
			6.0	MSG7260	GUIL20W47	2022-10	11	
			6.5	MSG7265	GUIL20W48	2022-10	11	
			7.0	MSG7270	GUIL20W48	2022-10	12	
			7.5	MSG7275	GUIL20W50	2022-11	198	
			8.0	MSG7280	GUIL20W51	2022-11	10	
			8.5	MSG7285	GUIL20W51	2022-11	10	
			9.0	MSG7290	GUIL20W51	2022-11	10	

Lot size as specified by client: 150,001 to 500,000 pieces

METHOD OF TEST:
1. EN 455-1:2020 Medical gloves for single use
   Part 1: Requirements and testing for freedom from holes
2. EN 455-2:2015 Medical gloves for single use
   Part 2: Requirements and testing for physical properties
   - Clause 4 Dimensions

Laboratory: TÜV SÜD PSB Pte. Ltd. TÜV SÜD @ IBP
Regional Head Office: TÜV SÜD Asia Pacific Pte. Ltd. TÜV SÜD @ IBP

Page 1 of 4



ARDL
Testing. Development. Problem Solving.


June 1, 2021
Zhiqiang Hu
Guilin HBM Health Protections, Inc.

Page 2 of 6
PN 159577

SUBJECT: Physical Testing on material submitted by the above company.

RECEIVED: Five (5) sets of surgical gloves identified as follows:
1) Medispo® PI by Guilin HBM Health Protections, Inc. Polyisoprene Sterile Surgical Gloves Powder Free
2) Protexis™ PI Surgical Gloves by Cardinal Health Sterile Polyisoprene Powder Free,
3) Gammex® Non-Latex PI Green by Ansell Healthcare Products LL Polyisoprene Surgical Gloves Powder Free
4) Biogel® UltraTouch by McKinlyckie Health Care AB Polyisoprene Surgical Gloves Powder Free
5) Sensicare® PI with Aloe Vera by Medline Industries Inc. Surgical Gloves Sterile Synthetic Polyisoprene Powder-Free with Aloe Vera

DECISION RULE: Rule #3



ARDL
Testing. Development. Problem Solving.

May 21, 2021

Guilin HBM Health Protections, Inc.

Page 2 of 7
PN 159121 – Rev 1

SUBJECT: Permeation testing per ASTM D6978-05(2019) on sample submitted by the above company.


RECEIVED: One (1) glove type identified by customer as: Medispo Polyisoprene Surgical Gloves, Lot No. HBM202105, Size 7.5.

TEST CHEMICALS:
Table 1: List of the Testing Drugs and their Sources

TESTING CHEMOTHERAPY DRUGS	DRUG SOURCE
Carboplatin, 10 mg/ml (10,000 ppm)	Teva, Lot# 19K11KA, Expiration 11/2021
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	USP, Lot# R116V0, Expiration 03/2022
Cisplatin, 1.0 mg/ml (1,000 ppm)	Accord, Lot# P200129S, Expiration 01/2022
Cyclophosphamide (Cytoxan), 20.0 mg/ml (20,000 ppm)	Accord, Lot# 19112225, Expiration 10/2021
Cytarabine HCL 100 mg/ml (100,000 ppm)	Sigma Aldrich, Lot# MKC-8892I, Expiration 01/2022
Dacarbazine, 10 mg/ml (10,000 ppm)	Teva, Lot# 313254148, Expiration 09/2021
Docetaxel, 10 mg/ml (10,000 ppm)	LC Labs, Lot# BDC-117, Expiration 01/2025
Doxorubicin HCl, 2.0 mg/ml (2,000 ppm)	WestWard, Lot# BJ0031, Expiration 06/2021
Etoposide, 20.0 mg/ml (20,000 ppm)	Teva, Lot# 313254660B, Expiration 07/2021
Fluorouracil, 50.0 mg/ml (50,000 ppm)	Accord, Lot# P200197, Expiration 01/2022
Fosfamide, 50 mg/ml (50,000 ppm)	Baxter Healthcare, Lot# 6AD19D, Expiration 01/2022
Mechlorethamine HCL 1 mg/ml (1,000 ppm)	Sigma Aldrich, Lot# MKBW4481V, Expiration 02/2022
Methotrexate, 25 mg/ml (25,000 ppm)	Teva, Lot# 19P09NB, Expiration 06/2021
Mitomycin C, 0.5 mg/ml (500 ppm)	USP, Lot# R07240, Expiration 10/2021
Mitoxantrone, 2 mg/ml (2,000 ppm)	USP, Lot# J0F27B, Expiration 01/2022
Paclitaxel, 6.0 mg/ml (6,000 ppm)	Teva, Lot# 19K24KA, Expiration 11/2021
Thiotepa, 10.0 mg/ml (10,000 ppm)	USP, Lot# R11580, Expiration 03/2022
Vinorelbine Bulfate, 1 mg/ml (1,000 ppm)	USP, Lot# Y06331, Expiration 05/2021

TESTING CONDITIONS:

Standard Test Method Used:	ASTM D6978-05(2019)
Analytical Method:	UV/VIS Spectrometry
Testing Temperature:	35.0°C ± 2.0
Collection System:	Closed Loop
Specimen Area Exposed:	5.067 cm²
Selected Data Points:	25/test
Number of Specimens Tested:	3/test
Location Sampled From:	Cuff



ARDL
Testing. Development. Problem Solving.

January 30, 2020
Ms. Cindy Xiao
Guilin HBM Health Protections, Inc.

Page 2 of 3
PN 152994

SUBJECT: Analytical testing on samples submitted by the above referenced customer

RECEIVED: One (1) glove type identified as MEDISPO Protein Reduced Sterile Latex Surgical Gloves

Decision Rule 1

ASSAY PROCEDURE:
Sample Extractions: The gloves were weighed and measured prior to extraction within a polypropylene extraction vessel. The extraction buffer used was 50 mM phosphate pH 7.4 at a ratio of 5 ml of buffer per gram of sample. The extraction was carried out at room temperature for two hours with agitation. The sponges were removed and the extract centrifuged at 500 xg for 15 minutes to pellet particulates. The cleared extract was then used in the assay.

ASTM D6712 Assay: Proteins are precipitated from the extract by the addition of a DOC-PTA-TCA (DeOxyCholate, PhosphoTungstic Acid, Tri-Chloroacetic Acid) solution and centrifugation. This precipitation step concentrates the proteins and removes substances that may interfere in the assay. The proteins are resuspended in 0.2 N NaOH and then transferred to each of two plates where they are incubated with a (1) copper tartrate solution and (2) an alkaline tartrate solution. Folin Ciocalteu's phenol reagent is added and incubated for 30 minutes for color to develop. The OD of the sample is then determined by reading at 700 nm. The OD values from the alkaline tartrate plate are subtracted from the copper tartrate plate. The resulting values are used to generate the Standard Curve. Protein values for the test items are determined by interpolation from this standard curve.

The LOD for the ASTM D5712 assay determined in the laboratory is 2.2 µg/ml and the LOQ is 11.0 µg/gm. An aqueous soluble protein content limit of 200 µg/dm² for gloves has been established. (See ASTM D3577 & ASTM D3578.)

RESULTS: The sample identified as MEDISPO Protein Reduced Sterile Latex Surgical Glove tested below the detection limit of the assay at ≤11.0 µg/gm and ≤8.6 µg/dm².

ASTM D5712 Test Certificate

Sample Description	Weight (gm)	Area (dm²)	Extract Vol (ml)	Assay Conc (µg/ml)		Total Protein (µg/gm)		Total Protein (µg/dm²)
MEDISPO Protein Reduced Sterile Latex Surgical Glove	9.9	10.6	48.5	b.d.				
	10.1	11.8	50.5	b.d.	<	11.0	<	9.6
	9.7	11.8	48.5	b.d.				

Where b.d. = below detection. (2.2 µg/ml)

*Note: The ASTM D5712 standard defines a "Sample" as three items. In the cases where less than three items are provided the Test Certificate is labeled as "Modified".



SATRA TECHNOLOGY

SATRA Technology Centre Ltd
Wyndham Way, Telford Way, Kettering,
Northamptonshire, NN16 8SD United Kingdom
Tel: +44 (0) 1536 410000
Fax +44 (0) 1536 410626
email: info@satra.com
www.satra.com

UKAS 0248

Customer details: Guilin HBM Health Protection, Inc.
No.1-2 Shujing East Road
Economic Development Zone
Guilin City
China

SATRA reference: CHM0318402/2134/JH/A
Your reference:
Date of report: 26th November 2021
Samples received: 28th October 2021
Date(s) work carried out: 2nd to 12th November 2021

TECHNICAL REPORT


Subject: EN 16523-1:2015+A1:2018 resistance to permeation by chemicals on gloves described as Medispo Polyisoprene Surgical gloves

Conditions of issue:
This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.
Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.
Tests marked # fall outside the UKAS Accreditation Schedule for SATRA.
All opinions and interpretations of results, and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.
A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.
Where values for uncertainty of measurement are included within the report then the uncertainty of the corresponding results are based on a standard uncertainty multiplied by a coverage factor k=2, which provides a coverage probability of approximately 95%.
When reporting results against a conformance statement (Pass/Fail or the allocation of a class or level) then uncertainty of measurement is taken into account based on a non-binary acceptance which itself is based on the guard band being equal to the expanded uncertainty.
Where the result corrected for uncertainty falls within the tolerance of the conformance statement then the risk of the conformance statement being a false accept or false reject is up to 2.5% and SATRA will in this instance quote a Pass/Fail, class or level.
Where the result corrected for uncertainty falls outside of the tolerance of the conformance statement then the risk of the conformance statement being a false accept or false reject is up to 50%. In this instance SATRA will not provide a Pass/Fail statement or a class or level but will include information in the notes in relation to the result obtained.
Please note that where uncertainty of measurement values have not been included then uncertainty has not been applied to these results. SATRA uncertainty of measurement values are however available upon request.

Report signed by: Jade Hurley
Position: Technologist
Department: Chemical & Analytical Technology

(Page 1 of 16)

SATRA Technology Centre Ltd (a subsidiary of SATRA). Registered in England No. 3992256 at the above address



Test Report No. 719124593S-EEC20/02-WBH
dated 06 Nov 2020

SUBJECT:
Testing of Gloves submitted by Guilin HBM Health Protections, Inc. on 12 Oct 2020.

TESTED FOR:
Guilin HBM Health Protections, Inc.
No.1-2, Shujing East Road,
Economic and Technological Development Area,
Guilin, China

TEST DATE:
13 Oct 2020 to 05 Nov 2020

DESCRIPTION OF SAMPLES:

S/N	Product Description	Brand	Colour	Size	Lot No.	Expiry Date	Sample received (pieces)	Manufacturer
1	Examination Gloves, Non Sterile, Latex, Powder Free	Medispo®	Natural	M	SRV20230821	202301	400	Guilin HBM Health Protections, Inc.

Lot size as specified by client: 150,001 to 500,000 pieces

METHOD OF TEST:
1. EN 455-1:2020 Medical gloves for single use
   Part 1: Requirements and testing for freedom from holes
2. EN 455-2:2015 Medical gloves for single use
   Part 2: Requirements and testing for physical properties
3. EN 455-3:2015 Medical gloves for single use
   art 3: Requirements and testing for biological evaluation

Page 1 of 5

Medispo®
Premium

POLYISOPRENE

SURGICAL GLOVES

Best Performance
Non-latex Synthetic





Medispo® PI Features:

Latex-free: Super soft polyisoprene formulation that is also tough to protect you and patients from pathogens. Exceeds all strength requirements according to USA and EU International standards.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium POLYISOPRENE SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Polyisoprene, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Latex-free, Polyisoprene
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body/Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Viral Penetration PPE Compliance (ASTM F1671), Chemotherapy Drugs (ASTM D6978), Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MPIG55	MPIG60	MPIG65	MPIG70	MPIG75	MPIG80	MPIG85	MPIG90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	290	290	290	290	290	290	290	300
Thickness(±0.03mm)	Fingertip: 0.23			Palm: 0.21			Cuff: 0.16	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308, www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

POLYISOPRENE

SURGICAL UNDERGLOVES

Best Performance
Non-latex Synthetic





Medispo® PI Under Features:

Perforation detection: Medispo PI Underglove is an ideal underglove when double-gloving. Designed to interlock with Medispo PI overglove to deliver an excellent single glove experience. Our distinct blue colour aids in clearly and quickly alerting wearers to perforations in the outer glove. Designed to be comfortable and reliable for any type of surgical procedure.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air-leaked and visually inspected.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium POLYISOPRENE UNDER SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Polyisoprene Under, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Latex-free, Polyisoprene
Color	Blue
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality / Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body / Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MPUG55	MPUG60	MPUG65	MPUG70	MPUG75	MPUG80	MPUG85	MPUG90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	290	290	290	290	290	290	290	300
Thickness(±0.03mm)	Fingertip: 0.21			Palm: 0.20			Cuff: 0.15	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

POLYISOPRENE

DOUBLE GLOVING SURGICAL GLOVES

Easy-don Double Gloving
Double Layer Protection





Medispo® PI Double Gloving Features:

Perforation detection: Medispo PI Double Gloving There are two types of gloves inside the double layer glove packaging: one pair of Medispo PI Sensitive outer gloves and one pair of ½ a size larger PI blue under gloves. Designed to be comfortable and reliable for any type of surgical procedure.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air leaked and visually inspected.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797

BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium POLYISOPRENE DOUBLE GLOVING SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® POLYISOPRENE DOUBLE GLOVING, Sterile, Hand Specific
Latex / Latex-free/ material	Latex-free, Polyisoprene
Color	White outer PI glove and blue under PI glove
Cuff type	Beaded cuff
Accelerators used	ZDEC (DPG Free)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile,E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65
International Product Standards	EN 455 1-4, EN ISO 374 1-5:2016, EN ISO 21420:2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485:2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745， USA 510k, MDSAP
Notified body/Test Laboratories	SATRA, ARDL, BSI
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residue

PRODUCT SPECIFICATION

Item No.	MPSD55	MPSD60	MPSD65	MPSD70	MPSD75	MPSD80	MPSD85
Size	5.5#+6.0#	6.0#+6.5#	6.5#+7.0#	7.0#+7.5#	7.5#+8.0#	8.0#+8.5#	8.5#+9.0#
Length(mm)	290	290	290	290	290	290	300
Thickness(±0.03mm)	Fingertip:0.21			Palm:0.20		Cuff:0.15	

PACKING INFORMATION

Pouch material	Long plastic peel pouch		
Packing Quantity	Pouch	Inner box	Shipping Carton
	2pairs/pouch	50pairs(25pouches)/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

POLYISOPRENE

SENSITIVE SURGICAL GLOVES

Designed For Super Sensitive Touch And Comfort





Medispo® PI Sensitive Features:

Sensitive tactility: Medispo PI Sensitive is designed for surgeons who demand a higher tactile sensitivity.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797

BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium POLYISOPRENE SENSITIVE SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Polyisoprene Sensitive, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Latex-free, Polyisoprene
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality / Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body / Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MPSE55	MPSE60	MPSE65	MPSE70	MPSE75	MPSE80	MPSE85	MPSE90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	290	290	290	290	290	290	290	300
Thickness(±0.03mm)	Fingertip: 0.21			Palm: 0.19			Cuff: 0.15	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

ORTHOPAEDIC

POLYISOPRENE SURGICAL GLOVES

Thicker And Stronger For Added Protection





Medispo® PI Orthopaedic Features:

Robust protection: Medispo PI Orthopaedic is our thickest glove in the PI portfolio. Our proprietary mould and microtextured finish delivers comfort with tactile sensitivity. Dark brown colour reduces glare from surgical lighting.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797

BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium

PI ORTHOPAEDIC SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Polyisoprene Orthopaedic, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Latex-free, Polyisoprene
Color	Brown
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality / Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body / Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MPOG55	MPOG60	MPOG65	MPOG70	MPOG75	MPOG80	MPOG85	MPOG90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	290	290	290	290	290	290	290	300
Thickness(±0.03mm)	Fingertip: 0.26			Palm: 0.24			Cuff: 0.18	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo® Premium

MEDISPO® TPC

LATEX SURGICAL GLOVES

Powder Free Latex
Our Best Value Product





Medispo® TPC Features:

Comfort, dexterity and protection: Building upon 20 years feedback from our customers HBM have designed a new mould to better fit with clinician's hands. Our new mould offers high-level comfort and dexterity. You may use Medispo TPC as your only glove, or as an overglove when double-gloving with a Medispo Latex Underglove to help protect you and your patients with perforation detection.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air-leaked and visually inspected.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797

BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

MEDISPO® TPC LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® TPC, Sterile, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex (NRL)
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body/Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Viral Penetration PPE Compliance (ASTM F1671), Chemotherapy Drugs (ASTM D6978), Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MTPC55	MTPC60	MTPC65	MTPC70	MTPC75	MTPC80	MTPC85	MTPC90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	280	280	280	285	285	290	290	300
Thickness(±0.03mm)	Fingertip: 0.22			Palm: 0.18			Cuff: 0.16	

PACKING INFORMATION

Pouch material	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308, www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

UNDERGLOVE

LATEX SURGICAL GLOVES

Easy-don Colored Undergloves
Perforation Detection System





Medispo® Latex Underglove Features:

Perforation detection: Medispo TPC Underglove is an ideal underglove when double-gloving. Designed to interlock with Medispo TPC overglove to deliver an excellent single glove experience. Our distinct green colour aids in alerting wearers to perforations in the outer glove. Designed to be comfortable and reliable for any type of surgical procedure.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air-leaked and visually inspected.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797

BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium

UNDERGLOVE LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Latex Underglove, Sterile, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex (NRL)
Color	Green
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body/Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MLUG55	MLUG60	MLUG65	MLUG70	MLUG75	MLUG80	MLUG85	MLUG90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	280	280	280	285	285	290	290	300
Thickness(±0.03mm)	Fingertip: 0.21			Palm: 0.20			Cuff: 0.14	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308, www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

DOUBLE GLOVING

LATEX SURGICAL GLOVES

Easy-don Double Gloving
Double Layer Protection



Medispo® Latex Double Gloving Features:

Perforation detection: Medispo Double Gloving There are two types of gloves inside the double layer glove packaging:one pair of Medispo Sensitive outer gloves and one pair of ½ a size larger green under gloves. Designed to be comfortable and reliable for any type of surgical procedure.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air leaked and visually inspected.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797

BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

DOUBLE GLOVING LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Latex DOUBLE GLOVING, Sterile, Hand Specific
Latex / Latex-free/ material	Natural rubber latex
Color	White outer glove and Green under glove
Cuff type	Beaded cuff
Accelerators used	ZDEC (DPG Free)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile,E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65
International Product Standards	EN 455 1-4, EN ISO 374 1-5:2016, EN ISO 21420:2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485:2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745， USA 510k, MDSAP
Notified body/Test Laboratories	SATRA, ARDL, BSI
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residue

PRODUCT SPECIFICATION

Item No.	MLSD55	MLSD60	MLSD65	MLSD70	MLSD75	MLSD80	MLSD85
Size	5.5#+6.0#	6.0#+6.5#	6.5#+7.0#	7.0#+7.5#	7.5#+8.0#	8.0#+8.5#	8.5#+9.0#
Length(±5mm)	280	280	280	285	285	290	290
Thickness(±0.03mm)	Fingertip:0.21			Palm:0.20		Cuff:0.14	

PACKING INFORMATION

Pouch material	Long plastic peel pouch		
Packing Quantity	Pouch	Inner box	Shipping Carton
	2pairs/pouch	50pairs(25pouches)/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

SENSITIVE

LATEX SURGICAL GLOVES

Designed For Super Sensitive Touch And Comfort





Medispo® Sensitive Features:

Sensitive tactility: Medispo Sensitive is designed for surgeons who prefer a natural rubber latex surgical glove but demand a higher tactile sensitivity.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

MEDISPO® SENSITIVE LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Sensitive, Sterile, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex (NRL)
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body/Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MLSG55	MLSG60	MLSG65	MLSG70	MLSG75	MLSG80	MLSG85	MLSG90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	280	280	280	285	285	290	290	300
Thickness(±0.03mm)	Fingertip: 0.18			Palm: 0.16			Cuff: 0.14	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308, www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

ORTHOPAEDIC

LATEX SURGICAL GLOVES

Thicker And Stronger For Added Protection





Medispo® Orthopaedic Features:

Robust protection: Medispo Orthopaedic is our thickest glove in the latex portfolio. Our proprietary mould and microtextured finish delivers comfort with tactile sensitivity. Warm brown colour reduces glare from surgical lighting.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium ORTHOPAEDIC LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Orthopaedic, Sterile, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex (NRL)
Color	Brown
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body/Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MLOG55	MLOG60	MLOG65	MLOG70	MLOG75	MLOG80	MLOG85	MLOG90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	280	280	280	285	285	290	290	300
Thickness(±0.03mm)	Fingertip: 0.30			Palm: 0.27			Cuff: 0.20	

PACKING INFORMATION

Pouch material and confi	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

PRISTINE™

LATEX SURGICAL GLOVES

Super-soft, Strong and Comfortable Ideal Companion For Medispo Latex Underglove



Medispo® PRISTINE™ Features:

Soft and tough: Medispo Pristine™ is our softest formulation glove, and delivers an enhanced level of comfort and tactility. The proprietary formulation of Pristine™ has been developed by our research and development team over several years and delivers a tough, robust glove with remarkable tensile strength and force at break.

Ideal double-glove companion: Medispo Pristine™ combined with Medispo green under-gloves offer an ideal combination for hole detection in the outer glove. Medispo double gloving system offers world leading protection and optimal comfort levels available. Medispo double gloving system is easy to don yet locks in place and behaves like one glove.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium

PRISTINE™ LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Pristine, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex (NRL)
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality / Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body / Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MPND55	MPND60	MPND65	MPND70	MPND75	MPND80	MPND85	MPND90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(±5mm)	280	280	280	285	285	290	290	300
Thickness(±0.03mm)	Fingertip: 0.22			Palm: 0.21			Cuff: 0.16	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

MEDISPO® POWDERED

LATEX SURGICAL GLOVES

Powdered Latex
Essential Product



Medispo® Powdered Features:

Comfort, dexterity and protection: One of our most economical glove offerings, you can rest assured Medispo Powdered will provide you and your patient excellent protection and comfort. Our new EUSA mould offers high-level comfort and dexterity. Medispo Powdered can be combined with Medispo Latex undergloves for additional protection with perforation detection capability.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air-leaked and visually inspected.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2696 | UDEM Adriatic d.o.o.

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

MEDISPO® POWDERED LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Powdered, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex (NRL)
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	sterile, E-Beam
Powder / Powder-free	Powdered
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	5 Years
Compliance to regulations	EU MDR 2017/745
Notified body/Test Laboratories	UDEM, SATRA, ARDL, TUV, SGS
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test

PRODUCT SPECIFICATION

Item No.	MLPG55	MLPG60	MLPG65	MLPG70	MLPG75	MLPG80	MLPG85	MLPG90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(Minimum)	250	260	260	270	270	270	280	280
Thickness(±0.03mm)	Fingertip: 0.2			Palm: 0.18			Cuff: 0.14	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308, www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

MEDISPO® Essential

LATEX SURGICAL GLOVES

Powder Free Latex
High performance-price ratio





Medispo® Essential Features:

Comfort, dexterity and protection: Building upon 20 years feedback from our customers HBM have designed a new mould to better fit with clinician's hands. Our new mould offers high-level comfort and dexterity. You may use Medispo Essential as your only glove, or as an overglove when double-gloving with a Medispo Latex Underglove to help protect you and your patients with perforation detection.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air-leaked and visually inspected.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

MEDISPO® Essential LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Essential, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex (NRL)
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	5 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body/Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MLES55	MLES60	MLES65	MLES70	MLES75	MLES80	MLES85	MLES90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(Minimum)	250	250	260	260	260	270	270	280
Thickness(±0.03mm)	Fingertip: 0.2			Palm: 0.18			Cuff: 0.14	

PACKING INFORMATION

Pouch material and configuration options	Full-length or square half-fold peel pouch or Paper-Paper packaging		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections


Medispo®
Premium

MEDISPO®

LATEX SURGICAL GLOVES

Powder Free Latex
High performance-price ratio





Medispo® Surgical Glove Features:

Comfort, dexterity and protection: Building upon 20 years feedback from our customers HBM have designed a new mould to better fit with clinician's hands. Our new mould offers high-level comfort and dexterity.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air-leaked and visually inspected.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: 510K, ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

MEDISPO® LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Surgical Glove, Sterile, Super-Soft, Ergonomic Design, EUSA Mould
Latex / Latex-free/ material	Natural Rubber Latex
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC; CAS No.14324-55-1
Accelerators NOT used	Does not contain: Diphhenylguanidine (DPG), Thiaurams, Mercaptobenzothaiazoles (MBT), P-phenylenediamine (PPD), Zinc dibutyldithiocarbamate (ZDBC), Dialkyl thioureas (DTU)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile, E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65 (Typically 0.30 – 0.45)
International Product Standards	EN 455 1-4, EN ISO 374 1-5: 2016, EN ISO 21420: 2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality / Environmental Standards	ISO 9001, ISO 13485: 2016, ISO 14001
Shelf Life	5 Years
Compliance to regulations	EU MDR 2017/745, USA 510k, MDSAP
Notified body / Test Laboratories	BSI, SATRA, ARDL, TUV, SGS
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Chemical Residuals test data available upon request

PRODUCT SPECIFICATION

Item No.	MLPF55	MLPF60	MLPF65	MLPF70	MLPF75	MLPF80	MLPF85	MLPF90
Size	5.5	6	6.5	7	7.5	8	8.5	9
Length(Minimum)	250	260	260	270	270	270	280	280
Thickness(±0.03mm)	Fingertip: 0.18			Palm: 0.16			Cuff: 0.13	

PACKING INFORMATION

Pouch material	Paper pouch		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

NEOPRENE

SURGICAL GLOVES

Best Performance
Non-latex Synthetic





Medispo® Neoprene Features:

Latex-free: Super soft Neoprene formulation that is also tough to protect you and patients from pathogens. Exceeds all strength requirements according to USA and EU International standards.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: EN 455 1-4, EN ISO 374 1-5

Compliance USA: ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium NEOPRENE SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Neoprene, Sterile, Hand Specific
Latex / Latex-free/ material	Latex-free, Neoprene
Color	Green
Cuff type	Beaded cuff
Accelerators used	ZDEC (DPG Free)
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile,E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65
International Product Standards	EN 455 1-4, ASTM D3577， AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485:2016, ISO 14001
Shelf Life	3 Years

PRODUCT SPECIFICATION

Item No.	MNEG-5.5	MNEG-6.0	MNEG-6.5	MNEG-7.0	MNEG-7.5	MNEG-8.0	MNEG-8.5	MNEG-9.0
Size	5.5#	6.0#	6.5#	7.0#	7.5#	8.0#	8.5#	9.0#
Length(±5mm)	290	290	290	290	290	290	290	300
Thickness(±0.03mm)	Fingertip:0.21			Palm:0.18			Cuff:0.15	

PHYSICAL PROPERTIES

	Before Aging	After Aging
Tensile Strength	17 Mpa, Min	12 Mpa, Min
Ultimate Elongation	Min 650 %	Min 490 %
Endotoxin	Max 500 EU/device	
Powder content	≤ 2 mg/glove	

PACKING INFORMATION

Pouch material	Long plastic peel pouch		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections

Medispo®
Premium

OBSTETRICS

LATEX SURGICAL GLOVES

Powder Free Latex
Gynecological Extended





Medispo® Obstetrics Features:

Comfort, dexterity and protection: Comfortable and more protective, designed specifically for gynecology: elongated sterilized gloves, suitable for gynecological and other surgeries, providing more protection. Comfortable and flexible.

Protection: Computer control and predictive quality analytics (AI) mean that each glove undergoes rigorous quality controls before release. In addition, each glove is 100% air-leaked and visually inspected.

Skin-friendly formulation: Our unique chlorine-free formulation production processes combined with the absence of chemical accelerators known to cause allergic responses mean skin irritations are minimised or eliminated. Tested according to ISO EN 10993.

Easy donning: Skin friendly polymer coated for easy donning even with damp hands saving you time.

Consistent calibrated grip: The outer surface of Medispo surgical gloves are specially treated to provide a consistent grip. The grip is monitored and controlled during production and each lot must pass stringent internal grip controls before being released.

CE 2797
BSI Group The Netherlands B.V.
Regulation (EU) 2017/745
BSI Group The Netherlands B.V.
Regulation (EU) 2016/425

Complies with the following Standards and Regulations:

Quality: ISO 9001, EN ISO 13485

Compliance EU: CE Mark, MDR, EN 455 1-4, EN ISO 374 1-5

Compliance USA: ASTM D3577

Environment: EN ISO 14001:2015

HBM Protections

Medispo® Premium OBSTETRICS LATEX SURGICAL GLOVES

PRODUCT INFORMATION

Product Name / Description	Medispo® Obstetrics, Sterile, Hand Specific
Latex / Latex-free/ material	Natural rubber latex
Color	White Cream
Cuff type	Beaded cuff
Accelerators used	ZDEC
Texture / Grip	Micro textured, calibrated grip, medium
Sterile / Non-Sterile / Method	Sterile,E-Beam
Powder / Powder-free	Powder-free
Design	Anatomical, forward thumb ball, hand specific
AQL value	0.65
International Product Standards	EN 455 1-4, EN ISO 374 1-5:2016, EN ISO 21420:2019, ASTM D3577, ASTM D6978, ASTM F739, ISO 10993, AS/NZS 4179, ISO 10282, JIS T9107
Quality/Environmental Standards	ISO 9001, ISO 13485:2016, ISO 14001
Shelf Life	3 Years
Compliance to regulations	EU MDR 2017/745， USA 510k, MDSAP
Notified body/Test Laboratories	SATRA, ARDL, BSI
PPE compliance / Type	PPE Regulation (EU) 2016/425, Type B
Biological Evaluation	EN 10993, Primary Skin Irritation Test, Dermal Sensitization Assay, Cytotoxicity Test
Additional tests	Viral Penetration PPE Compliance (ASTM F1671), Chemotherapy Drugs (ASTM D6978), Chemical Residue

PRODUCT SPECIFICATION

Item No.	MOSG55	MOSG60	MOSG65	MOSG70	MOSG75	MOSG80	MOSG85	MOSG90
Size	5.5#	6.0#	6.5#	7.0#	7.5#	8.0#	8.5#	9.0#
Length	⩾430	⩾430	⩾430	⩾430	⩾430	⩾430	⩾430	⩾430
Thickness(±0.03mm)	Fingertip:0.22			Palm:0.21			Cuff:0.16	

PACKING INFORMATION

Pouch material	Long or square plastic peel pouch;Paper pouch		
Packing Quantity	Pouch	Inner box	Shipping Carton
	1pair/pouch	50pairs/box	200pairs or 400pairs/carton

EUROPEAN REPRESENTATION

EC Rep & Registration No.	HBM Medical, Coliemore House, 3 Coliemore Road, Dalkey, Co Dublin, Ireland: 653382
EC Rep Contact Details	INFO@4SHBM.COM, Telephone +353 1 582 2308， www.4shbm.com
Manufacturer	Guilin HBM Health Protections Inc Address:No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin,China Emali:guilinhbm@hbmchina.com Web:www.hbmchina.com

HBM Protections



Technology, Quality & Precision
For A Greener Tomorrow
Our Planet, Our Future

Guilin HBM Health Protections Inc is situated in the beautiful environment of the Li River valley system.HBM is committed to protecting our local and global environments.

HBM Protections

Factory Info

Guilin HBM Health Protections, Inc.

No.1-2, Shuijing East Road, Economic and Technological Development Area, Guilin, China

guilinhbm@hbmchina.com

www.hbmchina.com

EC REP

HBM Medical

Coliemore House, Coliemore Road, Dalkey, Co Dublin, Ireland

info@4shbm.com

www.hbmmedical.com